



पाटन स्वास्थ्य विज्ञान प्रतिष्ठान Patan Academy of Health Sciences

(पाटन स्वास्थ्य विज्ञान प्रतिष्ठान एन, २०६४ अन्तर्गत स्थापित)
लगनखेल-५, ललितपुर, नेपाल Lagankhel -5, Lalitpur, Nepal

Ref. no: 'kha' 880-B/06/072/073

Date: June 07, 2016

To,

Ref: Revised Technical Specification of Supply Delivery and Installation of Hospital Equipments -PAHS/HE/16/072/073

This is writing to inform you that after having the discussion in pre-bid meeting dated on June 06, 2016, some Clauses of Technical Specifications have been amended in the bidding documents of Supply, Delivery and Installation of Hospital Equipments-PAHS/HE/16/072/073, those are approved by the authority and enclosed herewith.

Apart from the attached approved changes, all conditions of the aforementioned contract will be remaining the same. The interested bidders are requested to procure the bidding document and submit as per the amendments.

Thanking you.

Prof. Dr. Sangita Bhandary

Vice Chancellor, PAHS

Vice Chancellor

Supply Delivery and Installation of Hospital Equipments

Contract No: PAHS/HE/16/072/073

5. Infusion Pump

SN	Existing Technical Specification in Bidding Document	Modification
7.2	USFDA approved product certificate.	Should have USFDA or Japanese Standard and CE/ISO approved product certificate.

6. Syringe Pump

SN	Existing Technical Specification in Bidding Document	Modification
7.2	USFDA approved product certificate.	Should have USFDA or Japanese Standard and CE/ISO approved product certificate.

12. Ventilator

SN	Existing Technical Specification in Bidding Document	Modifications
4.4	Should have compressor as air source.	Should have compressor or built in compressor / turbine as air source.
4.15.6	PSV 0 -50 CM H2O	PSV 2 – 60cm H2O.
4.25	Upgradable to ETCO2.	ETCO2 feature to be removed.


Vice Chancellor


Shrestha